


AUG 23 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299 ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name Christopher M Babcock	ID# 3867 ✓ Date 8-19-2010 ✓
A Agency King Cove DPS Phone # 497-2210	
Instrument Location Squad Room	
B Alco S/N x124634 ✓ Target Value .078 ✓ High Pressure 500	
Alco Test Values .078 ✓ .078 ✓	
1st Alco 2nd Alco	
Signature 	
(OVER)	


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of Oct, 2010.

  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130299

BMB  
8/27/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130299

AUGUST 19, 2010

OPERATOR'S NAME:

BABCOCK CHRISTOPHER M

OPERATOR'S NUMBER: 3867

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 123456

DEPT/AGENCY: KUC1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .078

ALCO S/N: X124634

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.37 in  
ALCO TARGET .079 14:19  
BLANK TEST .000 14:20  
INTERNAL STANDARD VERIFIED 14:20  
ALCO TV 30.37 in .078 14:20  
BLANK TEST .000 14:21  
SUBJECT SAMPLE .000 14:21  
BLANK TEST .000 14:22  
ALCO TV 30.37 in .078 14:23  
BLANK TEST .000 14:24

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130299

AUGUST 19, 2010

TIME 14:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.37 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

1123456789:;<=>?@ABCDEF6  
HIJL#NOPQRSTU VWXYZ[\]^\_`abdefghijklmno  
pqrsuvwxyz{|}~